

ADVISORY NOTICE

18-2_2025_UREA 275

Dear customer,

We hereby inform you that reagent **UREA 275, LOT 2410005** (reg. n. 12000024, cat. n. XSYS0020) which is used for quantitative *in vitro* determination of Urea in human serum, plasma and urine, was found to show unacceptable results of calibrations (low value of blank and factor).

If you observe any issue with these batches of reagent during calibration or quality control measurement (such as blank/concentration/factor issues), please stop your analysis immediately, do not continue with other measurements with the same LOT (especially patient samples) and claim the product in our CRM system (support.erbamannheim.com). Once we receive your claim, we will provide you a replacement free of charge.

In case you have any additional question, please do not hesitate to contact us.

Best regards,

Erba Lachema s.r.o.

Erba Lachema s.r.o.
Karásek 2219/1d, 621 00 Brno
DIČ: CZ26918846 -7-



Brno, 22nd September 2025

Ref. No.AN_18-2_2025

