

## ADVISORY NOTICE

59\_2025\_H7100\_new software version SW A3.1.2

Dear customers,

We hereby inform you that the H7100 device software has been updated to SW version SW A3.1.2

**1. List of changes in Application Software ver. A3.1.2:**

- Improvements done: new algorithm of calculations for cells distribution is implemented to improve reading results for of low-level quality control.

For SW version A3.1.2, please contact your local service engineer or distributor.  
The latest version of the SW is available on the [ERBA Knowledge Base](#) website.

Applied for:

**INSTRUMENTS:**

50008197	INS00087	H7100 (220-240V) single set
50008485	INS00087	H7100 (115V) single set

In case you have any additional questions, please do not hesitate to contact us.

Best regards,  
Erba Lachema s.r.o.

**Erba Lachema s.r.o.**  
Karásek 2219/1d, 621 00 Brno  
DIČ: CZ26918846 -7-



Brno, 23<sup>th</sup> October 2025

Ref. No.AN\_59\_2025

