

## ADVISORY NOTICE

53\_2025\_CHEM 5v3\_New embedded software version v2025.C0.03

Dear customer,

We hereby inform you that the CHEM 5v3 device embedded software has been updated to SW version 2025.C0.03.

### 1. Purpose:

Embedded software v2025.C0.03 released to include new features, improvement and to address issues reported in previous application software ver.2025.C0.02.

### 2. List of changes in Embedded Software ver. 2025.C0.03:

- 1) New feature: In "MAINTENANCE" menu "ENABLE REAL TIME GRAPH" setting is added. After entering hidden key, "YES" or "NO" options will be available to set to enable or disable setting respectively
- 2) Bug fixed: In "MAINTENANCE" menu "KEY PAD TEST" menu item not found.

Embedded SW version **2025.C0.03** is already included in devices produced from SN S2647 onwards.

For SW versions, please contact your local service engineer or distributor.  
The latest version of the SW is available on the [ERBA Knowledge Base](#) website.

Applied for:

### INSTRUMENTS:

50007132	INS00013	CHEM 5v3
----------	----------	----------

In case you have any additional questions, please do not hesitate to contact us.

Best regards,  
Erba Lachema s.r.o.

**Erba Lachema s.r.o.**  
Karásek 2219/1d, 621 00 Brno  
DIČ: CZ26918846 -7-



Brno, 17<sup>th</sup> October 2025

Ref. No.AN\_53\_2025