

ADVISORY NOTICE

23_2026_Rheumatoid Factor

Dear customers,


We hereby inform you that a small number of individual vials within reagent **RF, LOT 2511033** (reg.n. 12000162, cat. n. XSYS0048) which are used for quantitative *in vitro* determination of Rheumatoid Factor in human serum, was found to show unacceptable results of calibration.

If you observe any issue with this batch of reagent during calibration or quality control measurement (Flag MONO or unacceptable control results), please stop your analysis immediately, do not continue with other measurements with the same LOT (especially patient samples) and claim the product in our CRM system (support.erbamannheim.com). Once we receive your claim, we will provide you a replacement free of charge.

In case you have any additional question, please do not hesitate to contact us.

Best regards,

Erba Lachema s.r.o.



Brno, 17th July 2026

Ref. No.AN_23_2026

