

ADVISORY NOTICE

20_2026_LAURA_ new software version sw 1.21

Dear customers,

We hereby inform you that the LAURA device software has been updated to SW version 1.21.

Purpose

As part of our continuous commitment to product improvement and measurement excellence, this software update introduces an enhancement to the evaluation algorithm used for MICROALBUPHAN LAURA test strips.

The update includes:

- Optimization of threshold settings for MICROALBUPHAN LAURA strip measurement
- Improved measurement robustness
- Reduced result variability
- Enhanced overall measurement consistency and user experience

These improvements further strengthen the performance and reliability of the analyzer during routine operation and support even more stable analytical evaluation.

SW version 1.21 is already included in devices produced from SN 2979-1-202605.

For SW version 1.21, please contact your local service engineer or distributor. The latest version of the SW is available on the [ERBA Knowledge Base](#) website.

Applied for:

INSTRUMENTS:

50005205	INS00063	LAURA
50005892	INS00063	LAURA

In case you have any additional questions, please do not hesitate to contact us.

Best regards,
Erba Lachema s.r.o.

Erba Lachema s.r.o.
Karásek 2219/1d, 621 00 Brno
DIČ: CZ26918846 -7-



Brno, 25th June 2026

Ref. No.AN_20_2026