

ADVISORY NOTICE

29_2026_Total Bilirubin, Change of Intended Use

Products: Bilirubin Total, BIL T 330 and BIL T 200
Ref. Code: XSYS0023 and BLT00010

Dear customers,

We want to inform you regarding to change of Intended use of Bilirubin Total kits BIL T 330 and BIL T 200. The assay is now also intended for monitoring bilirubin levels in neonates as an aid in the clinical management of neonatal hyperbilirubinemia.
Assay parameters and analytical performances were not changed.

Please update your instructions for use.

In case of any questions, please do not hesitate to contact your local ERBA team.

Best regards,

Erba Lachema s.r.o.
Technical Product Management



Brno, 31st August 2026

Ref. No.AN_29_2026