

TECHNICAL SPECIFICATIONS

ANALYTICAL MODES

Manual, Predilute, Capillary

TOTAL PARAMETERS

33: WBC, RBC, HGB, HCT, MCV, MCH, MCHC, RDW-CV, RDW-SD, PLT, MPV, PDW-CV, PDW-SD, PCT, P-LCR, P-LCC, Neu%, Lym%, Mon%, Eos%, Bas%, Neu#, Lym#, Mon#, Eos#, Bas#, ALY%*, LIC%*, ALY#*, LIC#*, cWBC#*, NRBC%*, NRBC#*

PRINCIPLE OF MEASUREMENT

RBC/PLT/WBC: Electrical Impedance DIFF: 3 Angle Laser Flow Cytomtery HGB: Cyanide Free Colorimetry MCV: Measured HCT: Calculated

GRAPHICS

3 Histograms (WBC/RBC/PLT) 4 Scatterplots (DIFF x3, BASO)

SAMPLE VOLUME

 $\label{eq:boost} \begin{array}{l} Whole \ Blood: 15\ \mu L\\ CBC \ Only: 11\ \mu L\\ Pre-diluted: 20\ \mu L\\ Capillary: 20\ \mu L \end{array}$

LINEARITY RANGE WBC (x 10^3/L): 0 - 300 RBC (x 10^6/L): 0.00 - 8.50

Hb (g/dL) : 0 - 25.0 HCT (%): 0 - 67 PLT (x 10^3/L) : 0 - 3000

CALIBRATOR

ELite H Cal (3mL) Open Vial Stability at 2-8°C: 7 Days

TRI-LEVEL CONTROLS Erba H5 CON L, N, H (3mL) Open Vial Stability at 2-8°C: 14 Days

THROUGHPUT 60 Tests/Hr

OC (L-J, X-BAR) Yes

> **DATA STORAGE** 50.000 Results with Graphs

DIMENSION (mm) 364 x 498 x 431

WEIGHT (Kg) 26.5

REAGENTS

Erba Dil (20L) ErbaLyse 1 (200mL) Erba Lyse 2 (500mL) Elite H Clean (50mL)

INTERFACES 4 USB + 1 LAN Port

OPERATING ENVIRONMENT 15-30°C Atmospheric pressure 70kPa~106kPa

POWER REQUIREMENT A.C.100-240V; 50/60 Hz; ≤200VA

* Research use only

H560 ORDER DETAILS

CAT. NO.	REG. NO.	PRODUCT NAME	DESCRIPTION
INS00078	50005216	H560	5 Part Differential

H-SERIES REAGENTS

CAT. NO.	REG. NO.	PRODUCT NAME	VOLUME
HEM00030	50005221	Erba H560 Diluent	20 Litres
HEM00031	50005222	Erba H560 Lyse1	200 mL
HEM00032	50005223	Erba H560 Lyse2	500 mL
HEM00023	50004878	ELite H Clean	50 mL
HEM00024	10020492	ELite H5 CON Low	3 mL
HEM00025	10020493	ELite H5 CON Normal	3 mL
HEM00026	10020494	ELite H5 CON High	3 mL
HEM00027	10020492	ELite H CAL	3 mL

EASY. EFFICIENT. RELIABLE.



Erba Lachema s.r.o. Karásek 2219/1d, 621 00 Brno, Czech Republic Phone: +420 517 077 111 E-mail: sales@erbamannheim.com www.erbamannheim.com

TOTAL SOLUTIONS FOR CLINICAL DIAGNOSTICS

 Image: Distribution of the state of the

lemətology Ana



H560 FULLY AUTOMATED 5 PART HEMATOLOGY ANALYSER

lirect ndirect sferase (ALT) sferase(AST hosphatase (serum) m) serum) ein eru

NAME

otal

timulating

e Protein

Upgrade to a five part (50 erur differential and report faster with more confidence.

1m/U4 033.

2,00 ng/dl

<4.0 mg

17.0

50.0 UII

31.0 UII

3.0 U S-H 88.0 µmolll 4.5 mmol/l

333.0 µmolll

68.0 gll

43.0 gll

4.40 mmolll


13.2 4.1


men: 4-37 UN en 18-120 years: 40-129 UN 61 UN SUS-H s: 53-97 µmol/l 8.3 mmol/l 50 gl normal: 3.5-5.5 mmol/l 21-1201 normal: 132-146 mmc normal: 99-109 mm 12-120 years: 2 5 0 mmol/l normal: 0.53


from 21

Three part differential systems are not designed to report results on abnormal patients without the need for further testing. These investigations are complex and delay the release of critical results to the requesting clinician.

The H560's five part differential count allows users to report world class results faster and with more confidence.








Blood film review is a complex process and takes time to master. Adding a five part differential system to the laboratory means that valuable time at the microscope can be spent assessing morphology and other abnormalities rather than counting.

With the H560 users can go beyond the normal 5 part differential with added information from the advanced research parameters - ALY%, LIC%, ALY#, LIC#, cWBC#, NRBC%, NRBC#.






Manual blood film review should focus on morphology, not counting.

H560 is ahead of the rest with class leading design and technologies.

Anti-clog technology

The impedance aperture is treated with a high energy pulse after each sample - reducing the risk of blockages





Detailed platelet information

The H560's PLCR and PLCC parameters allow users to report detailed information about the patient's platelet status



Reduce pre-analytical variables

To ensure consistent and exact dilution of the patient sample, an automated diluent dispense mode (ADDM)



with ADDM

has been created to help reduce errors



Wide range of tubes


The H560 is able to accept sample tubes from many different manufacturers including pediatric samples.

Just 15µL of Aspiration volum

The efficient fluidics design allows the H56 to aspirate only 15 µL. This combined with th open tube sampling aspiration means a coup of drops is more than enoug

Multiple analytical mode

Six analytical processing modes mean flexibility of analysis and ensure you get the best resu for your patient first time



15 µL	
ode & ID	
	CBC
Venous Whole Blood (VWB)	
Venous Whole Blood (VWB) Capillary Whole Blood (CWB)	Coo Dan
	00000

EASY **USER INTERFACE**

- 10.4" Touchscreen
- 50.000 Reports
- 4 Scatterplots (DIFF x3, BASO) • 3 Histograms
- (RBC, WBC, PLT) • 1-click analysis



ô O

RFIC

REAGENT REGISTRATION

• Full traceability via RFID inventory management system

RELIABLE **ENGINEERED CONSTRUCTION**



- Small footprint - 364 x 498 mm
- Status indicator
- Simple, well engineered construction
- High quality components for long life


ANTI-CLOG TECHNOLOGY

• Prevent build-up with anti-clog technology





EFFICIENT EXTERNAL CONNECTIONS



ANALYTICS



DESIGN

• 3 Angle flow differential results



• Low aspiration volume 15 µL • 33 parameters • Guaranteed dilution accuracy via automatic diluent dispensing • Advanced platelet analysis (P-LCR, P-LCC, PDW-SD, PDW-CV) • Automatic floating discriminations

cytometry for high quality WBC

